



CE Declaration of Conformity

We: Labconco Corporation
8811 Prospect Avenue
Kansas City, MO 64132 USA
Ph. 816-333-8811 Fax. 816-363-0130
labconco@labconco.com

being the representative party solely responsible for this certification, hereby declare that the following products:

Equipment: Laboratory Equipment – CentriVap Concentrators, Consoles & Systems

Type, Model: CentriVap DNA Concentrators 79700xx, where xx= 11, 30, 35, 37
CentriVap Benchtop Concentrators 78100xx, where xx= 30, 31, 32, 34, 35, 36, 38, 39, 40
CentriVap Acid Resistant Concentrators 78100xx, where xx= 17, 33, 37, 41
CentriVap Refrigerated Concentrators 73100xx, where xx= 30, 31, 32, 34, 35, 36, 37, 38, 39, 40, 41, 42
CentriVap Mobile Console Systems 78120xx, where xx= 11, 12, 14, 15, 35, 36, 37, 38
CentriVap Concentrator Systems
Aqueous Systems 79820xx, where xx= 11, 12, 35, 37
Acid Systems 79830xx, where xx= 13, 14, 15, 35, 37
Solvent Systems 7995800, where xx= 11, 12, 35, 37
Proteomic Systems 79896xx, where xx= 01, 02, 35, 37
CentriVap Complete 73150xx, where xx= 30, 31, 32, 33, 40, 41, 42, 43, 60, 61, 62, 63, 70, 71, 72, 73
CentriVap Cold Trap 78110xx, 74600xx, or 73850xx, where xx= 20, 21, 30, 31, 35, 36, 37, 38, 40, 41

Manufacturer: Labconco Corporation
8811 Prospect Avenue
Kansas City, MO 64132 USA
Ph. 816-333-8811 Fax. 816-363-0130
labconco@labconco.com

are in accordance with the following directives:

2014/30/EU Electromagnetic Compatibility Directive
2014/35/EU Low Voltage Directive
2006/42/EU Machinery Directive
2015/863/EU RoHS3 Directive

are in conformity with the standards listed below:

EN61010-1:2010	EN61000-4-3:2020
EN61326-1:2013	EN61000-4-4:2012
EN55011:2016+A1:2017+A11:2020	EN61000-4-5:2014+A1:2017
EN61000-3-2:2014	EN61000-4-6:2014
EN61000-3-3:2013	EN61000-4-11:2020
EN61000-4-2:2009	

when installed and operated in accordance with the manufacturers installation and operating instructions.

Brandon Gray
Vice President of Engineering
Labconco Corporation
Kansas City, MO USA

30 JUL 2021
Date of issue



UKCA Declaration of Conformity

We: Labconco Corporation
8811 Prospect Avenue
Kansas City, MO 64132 USA
Ph. 816-333-8811 Fax. 816-363-0130
labconco@labconco.com

being the representative party, in accordance with UK government guidance, solely responsible for this certification, hereby declare that the following product:

Equipment: Laboratory Equipment – CentriVap Concentrators, Consoles & Systems

Type, Model: CentriVap DNA Concentrators 79700xx, where xx= 11, 30, 35, 37
CentriVap Benchtop Concentrators 78100xx, where xx= 30, 31, 32, 34, 35, 36, 38, 39, 40
CentriVap Acid Resistant Concentrators 78100xx, where xx= 17, 33, 37, 41
CentriVap Refrigerated Concentrators 73100xx, where xx= 30, 31, 32, 34, 35, 36, 37, 38, 39, 40, 41, 42
CentriVap Mobile Console Systems 78120xx, where xx= 11, 12, 14, 15, 35, 36, 37, 38
CentriVap Concentrator Systems
Aqueous Systems 79820xx, where xx= 11, 12, 35, 37
Acid Systems 79830xx, where xx= 13, 14, 15, 35, 37
Solvent Systems 7995800, where xx= 11, 12, 35, 37
Proteomic Systems 79896xx, where xx= 01, 02, 35, 37
CentriVap Complete 73150xx, where xx= 30, 31, 32, 33, 40, 41, 42, 43, 60, 61, 62, 63, 70, 71, 72, 73
CentriVap Cold Trap 78110xx, 74600xx, or 73850xx, where xx= 20, 21, 30, 31, 35, 36, 37, 38, 40, 41

Manufacturer: Labconco Corporation
8811 Prospect Avenue
Kansas City, MO 64132 USA
Ph. 816-333-8811 Fax. 816-363-0130
labconco@labconco.com

the object of the declaration described above is in conformity with the relevant UK Statutory Instruments (and their amendments):

S.I. 2016/1091 Electromagnetic Compatibility Directive
S.I. 2016/1101 Low Voltage Directive
S.I. 2008/1597 Machinery Directive
S.I. 2012/3032 RoHS Directive

are in conformity with the standards listed below:

EN61010-1:2010	EN61000-4-3:020
EN61326-1:2013	EN61000-4-4:2012
EN55011:2016+A1:2017+A11:2020	EN61000-4-5:2014+A1:2017
EN61000-3-2:2014	EN61000-4-6:2014
EN61000-3-3:2013	EN61000-4-11:2020
EN61000-4-2:2009	

when installed and operated in accordance with the manufacturers installation and operating instructions.

Brandon Gray
Vice President of Engineering
Labconco Corporation
Kansas City, MO USA

30 JUL 2021
Date of issue