



# CE Declaration of Conformity

**We:** Labconco Corporation  
8811 Prospect Avenue  
Kansas City, MO 64132 USA  
Ph. 816-333-8811 Fax. 816-363-0130  
labconco@labconco.com

**Authorized Representative**  
Mauro Ferri c/o BIPIU'BI S.r.l.  
Via Vittorio Veneto, 12  
20091 Bresso (MI)  
Italia

*being the representative party solely responsible for this certification, hereby declare that the following products:*

Equipment: Laboratory Equipment – Logic (L4) Series Biological Safety Cabinets

Type, Model: 3023zxx-22, 32, 42, 52, 62, 72 Type A2 3' Wide; z=1 for 10" Sash, 2 for 12" Sash, 8 for 8" Sash  
3024zxx-22, 32, 42, 52, 62, 72 Type A2 4' Wide; z=1 for 10" Sash, 2 for 12" Sash, 8 for 8" Sash  
3025zxx-22, 32, 42, 52, 62, 72 Type A2 5' Wide; z=1 for 10" Sash, 2 for 12" Sash, 8 for 8" Sash  
3026zxx-22, 32, 42, 52, 62, 72 Type A2 6' Wide; z=1 for 10" Sash, 2 for 12" Sash, 8 for 8" Sash  
30348xx-22, 32, 42, 52, 62, 72 Type B2 4' Wide, 8" Sash  
30368xx-22, 32, 42, 52, 62, 72 Type B2 6' Wide, 8" Sash  
31232xx-22, 32, 42, 52, 62, 72 Type A2 Procedure Station 3' Wide 12" Sash  
31242xx-22, 32, 42, 52, 62, 72 Type A2 Procedure Station 4' Wide 12" Sash  
31252xx-22, 32, 42, 52, 62, 72 Type A2 Procedure Station 5' Wide 12" Sash  
31262xx-22, 32, 42, 52, 62, 72 Type A2 Procedure Station 6' Wide 12" Sash  
3y239xx-22, 32, 42, 52, 62, 72 Type A2 Cell Logic 3' Wide 9" Sash  
3y249xx-22, 32, 42, 52, 62, 72 Type A2 Cell Logic 4' Wide 9" Sash  
3y259xx-22, 32, 42, 52, 62, 72 Type A2 Cell Logic 5' Wide 9" Sash  
3y269xx-22, 32, 42, 52, 62, 72 Type A2 Cell Logic 6' Wide 9" Sash  
3y348xx-22, 32, 42, 52, 62, 72 Type B2 Cell Logic 4' Wide 8" Sash  
3y368xx-22, 32, 42, 52, 62, 72 Type B2 Cell Logic 6' Wide 8" Sash  
y = 2 for Temp Zone, 3 for Scope Ready, 4 for Temp Zone + Scope Ready  
x = any number 0 through 9

NOTE: Above model numbers may be followed by a 5-digit number which signify an approved modification. This modification does not impact the compliance to standards listed below.

Manufacturer: Labconco Corporation  
8811 Prospect Avenue  
Kansas City, MO 64132 USA  
Ph. 816-333-8811 Fax. 816-363-0130  
labconco@labconco.com

*are in accordance with the following directives:*

2014/30/EU	Electromagnetic Compatibility Directive	2011/65/EU	RoHS Directive
2014/35/EU	Low Voltage Directive	2015/863/EU	RoHS3 Directive
2006/42/EU	Machinery Directive		

*are in conformity with the standards listed below:*

EN61010-1:2010+A1:2019+A2:2021	EN61000-3-3:2013+A2:2021	EN61000-4-5:2014+A1:2017
EN61326-1:2021	EN61000-4-2:2009	EN61000-4-6:2022
EN55011:2016+A1:2017+A2:2021	EN61000-4-3:2020	EN61000-4-11:2020
EN61000-3-2:2019+A1:2021	EN61000-4-4:2012+A1:2022	EN63000:2018

*when installed and operated in accordance with the manufacturer's installation and operating instructions.*

**Brandon Gray**  
Vice President of Engineering  
Labconco Corporation  
Kansas City, MO USA

15 OCT 2024  
Date of issue



# UKCA Declaration of Conformity

**We:** Labconco Corporation  
8811 Prospect Avenue  
Kansas City, MO 64132 USA  
Ph. 816-333-8811 Fax. 816-363-0130  
labconco@labconco.com

**being the representative party, in accordance with UK government guidance, solely responsible for this certification, hereby declare that the following product:**

**Equipment:** Laboratory Equipment – Logic (L4) Series Biological Safety Cabinets

**Type, Model:**

3023zxx-22, 32, 42, 52, 62, 72	Type A2 3' Wide; z=1 for 10" Sash, 2 for 12" Sash, 8 for 8" Sash
3024zxx-22, 32, 42, 52, 62, 72	Type A2 4' Wide; z=1 for 10" Sash, 2 for 12" Sash, 8 for 8" Sash
3025zxx-22, 32, 42, 52, 62, 72	Type A2 5' Wide; z=1 for 10" Sash, 2 for 12" Sash, 8 for 8" Sash
3026zxx-22, 32, 42, 52, 62, 72	Type A2 6' Wide; z=1 for 10" Sash, 2 for 12" Sash, 8 for 8" Sash
30348xx-22, 32, 42, 52, 62, 72	Type B2 4' Wide, 8" Sash
30368xx-22, 32, 42, 52, 62, 72	Type B2 6' Wide, 8" Sash
31232xx-22, 32, 42, 52, 62, 72	Type A2 Procedure Station 3' Wide 12" Sash
31242xx-22, 32, 42, 52, 62, 72	Type A2 Procedure Station 4' Wide 12" Sash
31252xx-22, 32, 42, 52, 62, 72	Type A2 Procedure Station 5' Wide 12" Sash
31262xx-22, 32, 42, 52, 62, 72	Type A2 Procedure Station 6' Wide 12" Sash
3y239xx-22, 32, 42, 52, 62, 72	Type A2 Cell Logic 3' Wide 9" Sash
3y249xx-22, 32, 42, 52, 62, 72	Type A2 Cell Logic 4' Wide 9" Sash
3y259xx-22, 32, 42, 52, 62, 72	Type A2 Cell Logic 5' Wide 9" Sash
3y269xx-22, 32, 42, 52, 62, 72	Type A2 Cell Logic 6' Wide 9" Sash
3y348xx-22, 32, 42, 52, 62, 72	Type B2 Cell Logic 4' Wide 8" Sash
3y368xx-22, 32, 42, 52, 62, 72	Type B2 Cell Logic 6' Wide 8" Sash

y = 2 for Temp Zone, 3 for Scope Ready, 4 for Temp Zone + Scope Ready  
x = any number 0 through 9

NOTE: Above model numbers may be followed by a 5-digit number which signify an approved modification.  
This modification does not impact the compliance to standards listed below.

**Manufacturer:** Labconco Corporation  
8811 Prospect Avenue  
Kansas City, MO 64132 USA  
Ph. 816-333-8811 Fax. 816-363-0130  
labconco@labconco.com

**the object of the declaration described above is in conformity with the relevant UK Statutory Instruments (and their amendments):**

<b>S.I. 2016/1091</b>	<b>Electromagnetic Compatibility Directive</b>	<b>S.I. 2008/1597</b>	<b>Machinery Directive</b>
<b>S.I. 2016/1101</b>	<b>Low Voltage Directive</b>	<b>S.I. 2012/3032</b>	<b>RoHS Directive</b>

**are in conformity with the standards listed below:**

<b>EN61010-1:2010+A1:2019+A2:2021</b>	<b>EN61000-3-3:2013+A2:2021</b>	<b>EN61000-4-5:2014+A1:2017</b>
<b>EN61326-1:2021</b>	<b>EN61000-4-2:2009</b>	<b>EN61000-4-6:2022</b>
<b>EN55011:2016+A1:2017+A2:2021</b>	<b>EN61000-4-3:2020</b>	<b>EN61000-4-11:2020</b>
<b>EN61000-3-2:2019+A1:2021</b>	<b>EN61000-4-4:2012+A1:2022</b>	<b>EN63000:2018</b>

**when installed and operated in accordance with the manufacturer's installation and operating instructions.**

**Brandon Gray**  
Vice President of Engineering  
Labconco Corporation  
Kansas City, MO USA

15 OCT 2024  
Date of issue