


# Fumehood Performance Test Report

## 6ft Protector XStream Adverse Conditions Testing, 40 FPM

Prepared For

**Labconco Corporation**

Prepared By



***AccuTec***  
**Services, Inc.**  
[www.atsiusa.com](http://www.atsiusa.com)

**Report Date: 09-13-2012**

**Project No.: 121747**

## Table of Contents

<u>Title</u>	<u>Section</u>
<b>Summary</b>	<b>1</b>
<b>Modified NIH, 40 fpm</b>	
<b>Modified NIH, 40 fpm, with 50 fpm Cross Draft</b>	
<b>Modified NIH, 50 fpm, with Chest Height Gas     Detection</b>	
<b>Protector XStream Fume Hood Data</b>	<b>2</b>
<b>Certificates of Calibration</b>	<b>3</b>



## 1.0 Introduction

- 1.1 This report documents the work performed by AccuTec Services, Inc. and the test results for Labconco Corporation located in Kansas City, MO.
- 1.2 The scope of the project is to test the performance of Labconco Protector XStream Laboratory Fume Hoods in accordance with National Institute of Health (NIH) ASHRAE 110 with modifications as detailed below; to establish an "As Manufactured" rating for the equipment.

## 2.0 Summary

- 2.1 All test methods were performed in accordance with the guidelines established by the "On Site Testing Constant Volume Fume Hoods", 1999 Edition published by the National Institutes of Health, with variations noted in 2.2 – 2.4.
- 2.2 The Fume Hood was balanced to the design inlet velocity value; then tested to establish an "As Manufactured" performance rating and a velocity profile at 40 fpm. Test Data is included in Section 2 of this report.
- 2.3 The Fume Hood was tested with a 50 fpm Cross Draft to test adverse conditions.
- 2.4 The Fume Hood was tested at 50 fpm face velocity at chest height, to test concentrations of tracer gas directly across from the ejector.



### 3.0 Fume Hood Ratings

The XStream hood passed the acceptance criteria of 6.0AM-0.05 as listed in the specifications; individual ratings are listed in the table below.

40 FPM NIH Modified	
Fume Hood ID	Protector XStream
Rating	6.0AM-0.000



### 4.0 Abbreviations

4.1 The following abbreviations may have been used throughout this report and are represented here for reference.

SME	Sash Movement Effect	AM	As Manufactured
fpm	Feet per minute	Sec.	Seconds
Cal.	Calibration	L/m	Liters per minute
ppm.	Parts Per Million	No.	Number

## 5.0 Standards Utilized

- 5.1 The following calibrated field equipment was used in the execution of this work. Current certificates of calibration are provided in Section 3 of this report.

Device	Model	Serial No.
<b>TSI Anemometer</b>	966	P07460054
<b>Uson Leak Meter</b>	Q200	2000513
<b>BIOS Flow Meter</b>	Definer 220	119896



## Airflow Velocity Test Report

Project: Labconco  
Location: Manufacturer  
Manufacturer: Labconco  
Unit ID: --

Model: Protector XStream 2012  
Serial No.: --  
Type: Laboratory Fume Hood

Position of Sash (Percent Open)			<b>100%</b>
Average Airflow Face Velocity (fpm)			41
Highest Airflow Face Velocity (fpm)			56
Lowest Airflow Face Velocity (fpm)			27

<u>28"</u>	Actual Sash Height
<u>62.25"</u>	Actual Sash Width
<u>28"</u>	Design Sash Height
<u>62.25"</u>	Design Sash Width
<u>Vertical</u>	Sash Configuration
<u>40 fpm</u>	Design Airflow Face Velocity

Notes: 460 CFM at 0.07 In. W.C.


Acceptance Criteria:   X   Has Been Met        Has Not Been Met

Standards Utilized: P07460054 Cal. Due Date: 13-Aug-13


Technician: CLF Test Date: 25-Jul-12




## Airflow Velocity Test Report (Data)




Max Velocity:	55 fpm
Min Velocity:	31 fpm
Avg Velocity:	43 fpm




Max Velocity:	41 fpm
Min Velocity:	23 fpm
Avg Velocity:	31 fpm




Max Velocity:	41 fpm
Min Velocity:	27 fpm
Avg Velocity:	33 fpm




Max Velocity:	43 fpm
Min Velocity:	20 fpm
Avg Velocity:	30 fpm




Max Velocity:	51 fpm
Min Velocity:	33 fpm
Avg Velocity:	42 fpm




Max Velocity:	37 fpm
Min Velocity:	14 fpm
Avg Velocity:	27 fpm




Max Velocity:	72 fpm
Min Velocity:	33 fpm
Avg Velocity:	44 fpm




Max Velocity:	51 fpm
Min Velocity:	39 fpm
Avg Velocity:	44 fpm




Max Velocity:	51 fpm
Min Velocity:	31 fpm
Avg Velocity:	45 fpm




Max Velocity:	43 fpm
Min Velocity:	23 fpm
Avg Velocity:	36 fpm




Max Velocity:	45 fpm
Min Velocity:	33 fpm
Avg Velocity:	38 fpm




Max Velocity:	61 fpm
Min Velocity:	31 fpm
Avg Velocity:	43 fpm




Max Velocity:	61 fpm
Min Velocity:	21 fpm
Avg Velocity:	42 fpm




Max Velocity:	61 fpm
Min Velocity:	27 fpm
Avg Velocity:	44 fpm




Max Velocity:	61 fpm
Min Velocity:	31 fpm
Avg Velocity:	47 fpm



Max Velocity:	72 fpm
Min Velocity:	31 fpm
Avg Velocity:	51 fpm



Max Velocity:	55 fpm
Min Velocity:	35 fpm
Avg Velocity:	47 fpm



Max Velocity:	66 fpm
Min Velocity:	45 fpm
Avg Velocity:	56 fpm

Total:

Avg Velocity:	41 fpm
Max Velocity:	56 fpm
Min Velocity:	27 fpm

## Airflow Visualization Test Report (Local)

Project: Labconco  
Location: Manufacturer  
Manufacturer: Labconco  
Unit ID: --

Model: Protector XStream 2012  
Serial No.: --  
Type: Laboratory Fume Hood

Challenge:   X   Local        Large Volume

28" Sash Height  
62.25" Sash Width  
Vertical Sash Configuration  
Smoke Pen Challenge Medium

Narrative: Tested to NIH Protocol with simulated apparatus, consisting  
of two 3.8 L round paint cans, one 300 mm by 300 mm card-  
board box, and three 150 mm by 150 mm by 300 mm card-  
board boxes. These items are positioned randomly  
between 150 mm and 250 mm behind the sash and  
supported off the work surface by 50 mm by 50 mm blocks.

Airfoil Test:   X   Pass        Fail

Sash Perimeter Test:   X   Pass        Fail

Sash Bottom Test:   X   Pass        Fail

Work Surface Test:   X   Pass        Fail

Hood Cavity Test:   X   Pass        Fail

Sash Closed Perimeter/Interior Test:   X   Pass        Fail

Acceptance Criteria:   X   Has Been Met        Has Not Been Met

Notes: Tested to NIH protocol modification: 40 FPM average face velocity.

---

---

---

---

---

Technician: CLF

Test Date: 25-Jul-12

## Airflow Visualization Test Report (Large Volume)

Project: Labconco  
Location: Manufacturer  
Manufacturer: Labconco  
Unit ID: --

Model: Protector XStream 2012  
Serial No.: --  
Type: Laboratory Fume Hood

Challenge:      Local   X   Large Volume

28" Sash Height  
62.25" Sash Width  
Vertical Sash Configuration  
Glycol Fogger Challenge Medium

Narrative: Tested to NIH Protocol with simulated apparatus, consisting  
of two 3.8 L round paint cans, one 300 mm by 300 mm card-  
board box, and three 150 mm by 150 mm by 300 mm card-  
board boxes. These items are positioned randomly  
between 150 mm and 250 mm behind the sash and  
supported off the work surface by 50 mm by 50 mm blocks.

Airfoil Test:   X   Pass      Fail

Sash Perimeter Test:   X   Pass      Fail

Sash Bottom Test:   X   Pass      Fail

Work Surface Test:   X   Pass      Fail

Hood Cavity Test:   X   Pass      Fail

Sash Closed Perimeter/Interior Test:   X   Pass      Fail

Acceptance Criteria:   X   Has Been Met      Has Not Been Met

Notes: Tested to NIH protocol modification: 40 FPM average face velocity.

Technician: CLF Test Date: 25-Jul-12

## Tracer Gas Containment Test Report NIH Protocol at 40 FPM

Project:	Labconco	Model:	Protector XStream 2012
Location:	Manufacturer	Serial No.:	--
Manufacturer:	Labconco	Type:	Laboratory Fume Hood
Unit ID:	--		

	Right	Center	Left	Rapid Walk-by
Average Concentration:	0.000	0.000	0.000	0.000
Peak Concentration:	0.000	0.000	0.000	0.000

Static Mode Performance Rating: AM- 0.000


28"	Sash Height	Notes:	Tested to NIH protocol with simulated apparatus,
62.25"	Sash Width		consisting of two 3.8 L round paint cans, one 300 mm
Vertical	Sash Configuration		by 300 mm by 300 mm cardboard box, and three 150 mm,
6 L/m	Tracer Gas Release Rate		by 150 mm by 300 mm cardboard boxes. These items
			are positioned randomly between 150 mm to 250 mm
			behind the sash and supported off the work surface by
			50 mm by 50 mm blocks.

Acceptance Criteria:   X   Has Been Met        Has Not Been Met


Notes: Following right, center, and left tracer gas testing, a 5 minute rapid walk by test is conducted. Rapid walk by test has the manikin and ejector in the center position. Tracer gas levels are recorded while three rapid walk-by tests are conducted, at 300 mm behind the manikin, spaced 30 seconds apart.

Standards Utilized:	2000513	Cal. Due Date:	05-Jan-13
Technician:	CLF	Test Date:	25-Jul-12


**Tracer Gas Containment Test Report**  
**NIH Protocol at 40 FPM (Data)**



Average:	0.000	ppm
Peak:	0.000	ppm




Average:	0.000	ppm
Peak:	0.000	ppm



Average:	0.000	ppm
Peak:	0.000	ppm

**Tracer Gas Containment Test Report**  
**NIH Protocol at 40 FPM (Data)**



Average:	0.000	ppm
Peak:	0.000	ppm

## Tracer Gas Containment (Perimeter) Report

Project: Labconco  
Location: Manufacturer  
Manufacturer: Labconco  
Unit ID: --


Model: Protector XStream 2012  
Serial No.: --  
Type: Laboratory Fume Hood

### Location and Concentration of Leakage:

Indicate location of leaks and concentration below

<u>28"</u>	Sash Height	Notes: <u>Tested to NIH protocol modification: 40 FPM average face velocity.</u>
<u>62.25"</u>	Sash Width	
<u>Vertical</u>	Sash Configuration	
<u>6 L/m</u>	Tracer Gas Release Rate	

Perimeter leakage over time



Acceptance Criteria:   X   Has Been Met        Has Not Been Met

Standards Utilized: 2000513 Cal. Due Date: 05-Jan-13

Technician: CLF Test Date: 25-Jul-12

**Tracer Gas Containment Test Report  
NIH Protocol at 40 FPM Face Velocity  
with 50 fpm Cross Draft**

Project:	Labconco	Model:	Protector XStream 2012
Location:	Manufacturer	Serial No.:	--
Manufacturer:	Labconco	Type:	Laboratory Fume Hood
Unit ID:	--		

	Right	Center	Left
Average Concentration:	N/A	0.001	N/A
Peak Concentration:	N/A	0.054	N/A

Static Mode Performance Rating: AM- 0.001


28"	Sash Height	Notes:	Tested to NIH protocol with 50 FPM generated
62.25"	Sash Width		cross draft at 40 FPM face velocity.
Vertical	Sash Configuration		
6 L/m	Tracer Gas Release Rate		

Acceptance Criteria:   X   Has Been Met        Has Not Been Met

Standards Utilized:	2000513	Cal. Due Date:	05-Jan-13
Technician:	CLF	Test Date:	25-Jul-12



**Tracer Gas Containment Test Report  
NIH Protocol at 40 FPM Face Velocity  
with 50 fpm Cross Draft (Data)**



Average:	0.001	ppm
Peak:	0.054	ppm

**Tracer Gas Containment Test Report**  
**NIH Protocol**  
**50 FPM Face Velocity, Chest Detector**

Project:	Labconco	Model:	Protector XStream 2012
Location:	Manufacturer	Serial No.:	--
Manufacturer:	Labconco	Type:	Laboratory Fume Hood
Unit ID:	--		

	Right	Center	Left
Average Concentration:	0.000	0.000	0.000
Peak Concentration:	0.000	0.000	0.000

Static Mode Performance Rating: AM- 0.000


28"	Sash Height	Notes:	575 CFM at 0.10 In. W.C.
62.25"	Sash Width		Tested to NIH protocol with modifications: 50 FPM
Vertical	Sash Configuration		average face velocity. SF6 detector is the exact height of
6 L/m	Tracer Gas Release Rate		ejector, located in the chest of the manikin, 3" from the
			plane of the sash.

Acceptance Criteria:   X   Has Been Met        Has Not Been Met


Standards Utilized:                     2000513                     Cal. Due Date:             05-Jan-13            

Technician:                                 CLF                                 Test Date:             25-Jul-12


**Tracer Gas Containment Test Report**  
**NIH Protocol**  
**50 FPM Face Velocity, Chest Detector**



Average:	0.000	ppm
Peak:	0.000	ppm



Average:	0.000	ppm
Peak:	0.000	ppm



Average:	0.000	ppm
Peak:	0.000	ppm





# Bios

Driving a Higher Standard  
in Flow Measurement<sup>SM</sup>

## Calibration Certificate

Certificate No. 5010780

Product Definer 220 Medium Flow

Serial No. 119896

Cal. Date 12/19/2011

Sold to: AccuTec Services Inc. - Lee's Summit

31211 East Colbern Road

Grain Valley, MO 64029

USA

All calibrations are performed in accordance with ISO 17025 at Bios International Corporation, 10 Park Place, Butler, NJ, 07405, 800-663-4977, an ISO 17025:2005 – accredited laboratory through NVLAP. This report shall not be reproduced except in full without the written approval of the laboratory. Results only relate to the items calibrated. This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the Federal Government.

All units tested in accordance with Bios International Corporation test number PR18-13 using high-purity bottled nitrogen or dry filtered laboratory air.

### As Received Calibration Data

Technician Sonia Otero

Lab. Pressure 766 mmHg

Lab. Temperature 22.4 °C

Instrument Reading	Lab Standard Reading	Deviation	Allowable Deviation	As Received
100.65 sccm	100.875 sccm	-0.22%	1.00%	In Tolerance
1003.9 sccm	1005.95 sccm	-0.2%	1.00%	In Tolerance
5016.8 sccm	5007.95 sccm	0.18%	1.00%	In Tolerance
22.8 °C	22.4 °C	0.4°C	±0.8°C	In Tolerance
765 mmHg	766 mmHg	-1 mmHg	±3.5mmHg	In Tolerance

### Bios International Standards Used

Description	Standard Serial Number	Calibration Date	Calibration Due Date
ML-800-24	117991	3/29/2011	3/28/2012
Precision Thermometer	305460	8/15/2011	8/14/2012
Precision Barometer	2981392	5/27/2011	5/26/2012



# Bios

Driving a Higher Standard  
in Flow Measurement<sup>SM</sup>

## As Shipped Calibration Data

Certificate No. 5010780  
Technician Sonia Otero

Lab. Pressure 754 mmHg  
Lab. Temperature 22.4 °C

Instrument Reading	Lab Standard Reading	Deviation	Allowable Deviation	As Shipped
100.59 sccm	100.125 sccm	0.46%	1.00%	In Tolerance
1005.3 sccm	1001.65 sccm	0.36%	1.00%	In Tolerance
5032.4 sccm	5006.9 sccm	0.51%	1.00%	In Tolerance
22.3 °C	22.3 °C	-	±0.8°C	In Tolerance
754 mmHg	754 mmHg	-	±3.5mmHg	In Tolerance


## Bios International Standards Used

Description	Standard Serial Number	Calibration Date	Calibration Due Date
ML-800-24	100439	4/21/2011	4/20/2012
Precision Thermometer	305460	8/15/2011	8/14/2012
Precision Barometer	2981392	5/27/2011	5/26/2012

### Calibration Notes

Bios is an ISO 17025-accredited metrology laboratory. Each Bios primary gas flow standard is dynamically verified by comparing it to one of our laboratory standards, which is a Proven DryCal® Technology volumetric piston prover of much higher accuracy but of similar operating principles. For this purpose, a flow generator of ±0.03% stability is used. Our laboratory standards are qualified by direct measurement of their dimensions (diameter, length and time) using NIST-traceable precision gauges and instruments, such as depth micrometers and laser micrometers. NIST numbers for these gauges and instruments are available upon request. Rigorous analyses of our laboratory standards' uncertainties have been performed, in accordance with The Guide to the Expression of Uncertainty in Measurement (the GUM), assuring their traceable accuracy. Flow readings in sccm performed at STP of 21.1°C and 760 mmHg.

### Technician Notes:



Harvey Padden, President and Chief Metrologist



# CERTIFICATE OF CALIBRATION AND TESTING

TSI Incorporated, 500 Cardigan Road, Shoreview, MN 55126 USA  
Tel: 1-800-874-2811 1-651-490-2811 Fax: 1-651-490-3824 <http://www.tsi.com>

ENVIRONMENT CONDITION			MODEL	966
TEMPERATURE	69.6 (20.9)	°F (°C)	SERIAL NUMBER	P07460054
RELATIVE HUMIDITY	49	%RH		
BAROMETRIC PRESSURE	28.95 (980.4)	inHg (hPa)		

☒ AS LEFT  
☐ AS FOUND

☒ IN TOLERANCE  
☐ OUT OF TOLERANCE

## - CALIBRATION VERIFICATION RESULTS -

HUMIDITY VERIFICATION				SYSTEM H-102				Unit: %RH
#	STANDARD	MEASURED	ALLOWABLE RANGE	#	STANDARD	MEASURED	ALLOWABLE RANGE	
1	10.0	12.7	7.0~13.0	4	70.0	70.2	67.0~73.0	
2	30.0	31.6	27.0~33.0	5	90.0	89.2	87.0~93.0	
3	50.0	51.0	47.0~53.0					

TEMPERATURE VERIFICATION				SYSTEM T-101				Unit: °F (°C)
#	STANDARD	MEASURED	ALLOWABLE RANGE	#	STANDARD	MEASURED	ALLOWABLE RANGE	
1	32.0 (0.0)	32.0 (0.0)	31.5~32.5 (-0.3~0.3)	2	140.0 (60.0)	140.0 (60.0)	139.5~140.5 (59.7~60.3)	

VELOCITY VERIFICATION				SYSTEM V-111				Unit: ft/min ( m/s )
#	STANDARD	MEASURED	ALLOWABLE RANGE	#	STANDARD	MEASURED	ALLOWABLE RANGE	
1	0 (0.00)	0 (0.00)	-3~3 (-0.02~0.02)	7	644 (3.27)	647 (3.29)	625~664 (3.17~3.37)	
2	35 (0.18)	36 (0.18)	32~38 (0.16~0.20)	8	999 (5.08)	1001 (5.09)	969~1029 (4.92~5.23)	
3	65 (0.33)	65 (0.33)	62~68 (0.32~0.35)	9	1477 (7.50)	1493 (7.58)	1432~1521 (7.28~7.73)	
4	100 (0.51)	100 (0.51)	97~103 (0.49~0.53)	10	2493 (12.66)	2500 (12.70)	2418~2568 (12.28~13.04)	
5	161 (0.82)	161 (0.82)	156~165 (0.79~0.84)	11	4511 (22.92)	4504 (22.88)	4376~4647 (22.23~23.61)	
6	325 (1.65)	321 (1.63)	315~335 (1.60~1.70)	12	7985 (40.56)	8028 (40.78)	7746~8225 (39.35~41.78)	

TSI does hereby certify that the above described instrument conforms to the original manufacturer's specification (not applicable to As Found data) and has been calibrated using standards whose accuracies are traceable to the United States National Institute of Standards and Technology (NIST) or has been verified with respect to instrumentation whose accuracy is traceable to NIST, or is derived from accepted values of physical constants. TSI's calibration system is registered to ISO-9001:2008 and meets the requirements of ISO 10012:2003.

Measurement Variable	System ID	Last Cal.	Cal. Due
Barometric Pressure	E001992	04-06-12	04-06-13
Temperature	E004398	06-05-12	12-05-12
Pressure	E001058	01-18-12	01-18-13
Temperature	E003986	04-17-12	10-17-12
Humidity	E003539	02-28-12	08-28-12

Measurement Variable	System ID	Last Cal.	Cal. Due
DC Voltage	E004018	07-12-12	07-12-13
Pressure	E004041	03-30-12	09-30-12
Velocity	E003327	09-19-07	09-19-12
Temperature	E003987	04-17-12	10-17-12

*Callmeier*  
CALIBRATED

August 14, 2012

DATE

Doc ID: CERT\_GEN\_WCC

## MODEL 200 CERTIFICATE OF CALIBRATION

Certificate Number: 200-9392	Customer Order No:
Customer: Labconco Corporation	Service report No 55458
Model Number: Q200	Serial Number : 2000513

### CALIBRATION PROCEDURE

- The instrument was switched on for a period of 30 minutes prior to calibration.
- The instrument was calibrated at the CALIBRATION point to give a correct display.
- The calibration was carried out using the following devices that are traceable to national standards.

#### 1. DEVICES USED

Temperature Indicator Ai 0759

Ambient temperature during calibration 22.8 °C

#### 2. CALIBRATION POINT & SINTER TYPE FITTED AT CALIBRATION

Cal @ 1x10<sup>-6</sup> ml/s **H** ☒

Cal @ 1x10<sup>-6</sup> ml/s **M** ☐

Cal @ 1x10<sup>-5</sup> ml/s **L** ☐

#### 3. SF<sub>6</sub> STANDARD LEAK SERIAL NUMBERS & RESULTS

SF <sub>6</sub> Applied Sample ml/s (ppm)	Standard Leak (Serial Number)	Pre Adjustment Value* (ppm)	Post Adjustment Value* (ppm)
0.01± 0.01	1.0 x 10 <sup>-8</sup> (26312)	0.00	0.01 ppm
0.10± 0.01	1.0 x 10 <sup>-7</sup> (28185)	0.05	0.10 ppm
1.00± 0.1	1.0 x 10 <sup>-6</sup> (27459)	0.44	1.00 ppm
10.0± 1.0	1.0 x 10 <sup>-5</sup> (29655)	4.56	10.18 ppm


Terms used, H, M, L Sinter type, ml/s milli-litres per second, Std standard, N/A - Not Applicable, SF<sub>6</sub> - Sulphur Hexafluoride, PPM - part per million.

\* Indication relative to atmospheric concentration

#### 4. APPROVAL

Pre-adjustment Calibration	Passed	<input type="checkbox"/>	Failed	<input checked="" type="checkbox"/>	Refer to service report for failure information.
Post-adjustment Calibration	Passed	<input checked="" type="checkbox"/>	Failed	<input type="checkbox"/>	

Calibrated by N.P.Elbourne



Date of Calibration

5<sup>th</sup> January 2012